



FDA Session

AI Forum
February 9, 2024

<https://www.mcra.com/>

Hi, My Name is...



Alex Cadotte

Sr. Director, Digital Health & Radiology Regulatory Affairs

Previously employed by FDA as Lead Reviewer and Team Lead.

acadotte@mcra.com




THE LEADING MEDICAL DEVICE, DIAGNOSTICS & BIOLOGICS
CRO WITH INTEGRATED CLINICAL, REGULATORY,
REIMBURSEMENT, COMPLIANCE, & QUALITY

Our Integrated Services & Offerings

 CLINICAL RESEARCH ORGANIZATION (CRO)	 US REGULATORY	 INTERNATIONAL REGULATORY	 REIMBURSEMENT, HEALTH ECONOMICS & MARKET ACCESS	 QUALITY ASSURANCE	 HEALTHCARE COMPLIANCE	 CYBERSECURITY
--	--	--	--	--	---	--

MCRA's Integrated Advisory Service Lines



Today's Objectives

- Answer the following questions:
 - Is your technology regulated by FDA?
 - What testing will you need to perform to gain FDA clearance or approval?
 - How is the testing done? Nima - MCRA's AI & Imaging Center
 - What pathways are available to you?
 - What should you plan to do next?

A photograph of the U.S. Capitol building at night, illuminated by warm lights. A large white circle is superimposed over the center of the image, framing the text. The building's dome and classical architecture are clearly visible against the dark sky.

Regulatory 101

Statutory Definition of Medical Device

Per Section 201(h) of the Food, Drug, and Cosmetic Act, a device is:

An instrument, apparatus, implement, machine, contrivance, implant, in vitro reagent, or other similar or related article, including a component part, or accessory which is:

- recognized in the official National Formulary, or the United States Pharmacopoeia, or any supplement to them,*
- intended for use in the diagnosis of disease or other conditions, or in the cure, mitigation, treatment, or prevention of disease, in man or other animals, or*
- intended to affect the structure or any function of the body of man or other animals, and which does not achieve its primary intended purposes through chemical action within or on the body of man or other animals and*

which does not achieve its primary intended purposes through chemical action within or on the body of man or other animals and which is not dependent upon being metabolized for the achievement of its primary intended purposes. The term "device" does not include software functions excluded pursuant to section 520(o).


Intended Use vs. Indications for Use

Intended Use

refers to objective intent which can be shown by:

- Labeling
- Advertising materials
- Oral/written statements
- Implied or expressed claims

21 CFR 801.4




Use Case and Labeling Claims

- Let's craft an indications for use statement:
 - Who are your Target User(s)?
 - What disease or population are look targeting?
 - Where do you intend/expect your product to be used?
 - How do you want your device or device outputs to be used?
 - Do you expect to sell your product direct to patient/consumer or through a physician?

The indications for use statement is the foundation for all activities that involve the FDA

Classify Your Device



Class I

General Controls (with and without exemptions)

Most Class I devices are exempt from premarket notification process (510(k))



Class II

General and Special Controls (with and without exemptions)

Most Class II devices are without exemptions and require either a De Novo or 510(k) submission.



Class III

General Controls and Premarket Approval

All Class III devices are required to submit a premarket approval (PMA) application

Does Your Product Include Software?

21st Century Cures Act established exclusions for certain software functions from the definition of a device, as defined in section 201(h) of the FD&C Act.

- + Administrative functions
- + Low-risk general wellness
- + Electronic patient records

- + Medical device data systems
- + Clinical decision support software



Identify Applicable Regulation and Procode

The image displays three overlapping screenshots of the U.S. Food & Drug Administration (FDA) website, specifically the Medical Devices section. The top-left screenshot shows the 'Premarket Approval (PMA)' page, which includes a search form with fields for Applicant, Product Code, PMA Number, Device, Expedited Review, Decision Date, Docket Number, Advisory Committee, and Supplement Type. The top-right screenshot shows the '510(k) Premarket Notification' page, featuring a search form with fields for 510K Number, Type, Product Code, Center, Applicant Name, Device Name, and checkboxes for Combination Products, Cleared/Approved In Vitro Products, Redacted FOIA 510(k), Third Party Reviewed, and Clinical Trials. The bottom screenshot shows the 'Device Classification Under Section 513(f)(2)(De Novo)' page, which includes a search form with fields for Denovo Number, 510(k) Number, Panel, Center, Device Name, Requester Name, Decision Date, and Sort By. All three pages include a 'Search Database' button and a 'Quick Search' link. The FDA logo and navigation menu are visible at the top of each page. A sidebar on the right of each page lists 'Other Databases' such as De Novo, Medical Device Reports (MAUDE), CDRH Export Certificate Validation (CECV), CDRH FOIA Electronic Reading Room, CFR Title 21, CLIA, Device Classification, FDA Guidance Documents, Humanitarian Device Exemption, Medsun Reports, Post-Approval Studies, Postmarket Surveillance Studies, Radiation-Emitting Products, Radiation-Emitting Electronic Products Corrective Actions, Recalls, Registration & Listing, Standards, Total Product Life Cycle, and X-Ray Assembler.

Premarket Approval (PMA)

Premarket approval (PMA) is the FDA process of scientific and regulatory review to evaluate the safety and effectiveness of Class III medical devices. Class III devices are those that support or sustain human life, are of substantial importance in preventing impairment of human health, or which present a potential, unreasonable risk of illness or injury.

510(k) Premarket Notification

A 510(k) is a premarket submission made to FDA to demonstrate that the device to be marketed is as safe and effective, that is, substantially equivalent, to a legally marketed device (section 513(i)(1)(A) FD&C Act) that is not subject to premarket approval.

Device Classification Under Section 513(f)(2)(De Novo)

In 1997, the Food and Drug Administration Modernization Act (FDAMA) added the De Novo classification pathway under Section 513(f)(2) of the FD&C act, establishing an alternate pathway to classify new devices into class I or II that had automatically been placed in class III after receiving a Not Substantially Equivalent (NSE) determination in response to a 510(k) submission. In this process, a sponsor who receives an NSE determination may, within 30 days of receiving notice of the NSE determination, request FDA to make a risk-based classification of the device under section 513(a)(1) of the act.

In 2012, section 513(f)(2) of the FD&C act was amended by section 607 of the Food and Drug Administration Safety and Innovation Act (FDASIA), to provide a second option for De Novo Classification. In this second pathway, a sponsor who determines that there is no legally marketed device upon which to base a determination of Substantial Equivalency may request FDA to make a risk-based classification of the device under section 513(a)(1) of the act without first submitting a 510(k).

Where to start

1. Do you have any competitor products?
2. Search for product classification with general device types

<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>

Still Unsure? Submit a 513(g)

Contains Nonbinding Recommendations

FDA and Industry Procedures for Section 513(g) Requests for Information under the Federal Food, Drug, and Cosmetic Act

Guidance for Industry and Food and Drug Administration Staff

Document issued on December 16, 2019.

Document originally issued on December 21, 2015.

**This document supersedes, FDA and Industry Procedures for Section 513(g)
Requests for Information under the Federal Food, Drug, and Cosmetic Act,
issued December 21, 2015.**

Let's Pause to Ask Ourselves...

Is your technology or feature(s) regulated by FDA?

If yes, is it Class I, II, or III?

A photograph of the U.S. Capitol building at night, illuminated by warm lights. A large white circle is superimposed over the center of the image, framing the text. The text "Pre-Market Requirements" is written in a bold, white, sans-serif font across the middle of the circle.

Pre-Market Requirements

Valid Scientific Evidence

- Where do I go to identify what I need to do:
 - Previously cleared and approved decisions summaries
 - FDA Guidance Documents
 - FDA Recognized Consensus Standards
- Evidence to be Considered:
 - Bench Top V&V
 - Software Validation
 - Cybersecurity Testing
 - Biocompatibility
 - Sterilization and Packaging Testing
 - Electrical Safety
 - Electromagnetic Compatibility
 - Wireless Performance
 - Human Factors & Usability Testing
 - Clinical Validation

Determining the Evidence Needed

- Critical questions to ask that may eliminate evidence needed:
 - Does my product include hardware?
 - Does my product include AI/ML algorithms?
 - Does my product use off-the-shelf components/devices?
 - Does my product include in-house developed software?
 - Does my product make physical contact with a patient/user?
 - Does my product enter a sterile environment?
 - Does my product share information/data wirelessly?
 - Does my product connect to another device?

Clinical Validation Testing

III. SIGNIFICANT RISK AND NON-SIGNIFICANT RISK DEVICE STUDIES

A. What is a Significant Risk Device Study?

Under 21 CFR 812.3(m), an SR device means an investigational device that:

- Is intended as an implant and presents a potential for serious risk to the health, safety, or welfare of a subject;
- Is purported or represented to be for use supporting or sustaining human life and presents a potential for serious risk to the health, safety, or welfare of a subject;
- Is for a use of substantial importance in diagnosing, curing, mitigating, or treating disease, or otherwise preventing impairment of human health and presents a potential for serious risk to the health, safety, or welfare of a subject; or
- Otherwise presents a potential for serious risk to the health, safety, or welfare of a subject.

B. What is a Nonsignificant Risk Device Study?

An NSR device study is one that does not meet the definition for an SR device study.

C. Who Decides Whether A Device Study is SR or NSR?

Sponsors are responsible for making the initial risk determination and presenting it to the IRB. FDA is also available to help the sponsor, clinical investigator, and IRB in making the risk determination.²

Investigational Device Exemption (IDE) is required. Full compliance with 21 CFR 812 is required.

Investigational Device Exemption (IDE) is **NOT** required. Partial compliance with 21 CFR 812 is required.

<https://www.fda.gov/media/75459/download>

10 Guiding Principles for AI/ML Devices

- Does your product include AI/ML-based software?

1. Multi-Disciplinary Expertise Is Leveraged Throughout the Total Product Life Cycle
2. Good Software Engineering and Security Practices are Implemented
3. Clinical Study Participants and Data Sets Are Representative of the Intended Patient Population
4. Training Data Sets are Independent of Test Sets
5. Selected Reference Data Sets are Based Upon Best Available Methods
6. Model Design is Tailored to the Available Data and Reflects the Intended Use of the Device
7. Focus is Placed on the Performance of Human-AI Team
8. Testing Demonstrates Device Performance During Clinically Relevant Conditions
9. Users are Provided Clear, Essential Information
10. Deployed Models are Monitored for Performance and Retraining Risks are Managed

Class III PMA Pre-Approval Inspection

- Class II FDAs and De Novos are not subject to pre-authorization inspection by FDA but are subject to post-market inspection.
- ALL Class III require onsite inspection BEFORE PMA can be approved, called the QSIT (or quality system inspection technique):
 - Management controls
 - Design Controls
 - Corrective and Preventative Actions
 - Production and Process Controls
 - Sampling Plans

<https://www.fda.gov/files/Guide-to-Inspections-of-Quality-Systems.pdf>


Let's Pause to Ask Ourselves...

What testing do you think you need to support a marketing application?

A photograph of the U.S. Capitol building at night, with its iconic dome and illuminated windows. The image is overlaid with a dark purple gradient and a large white circle that frames the central text.


FDA Submission Types and Timelines

Pre-Submission Timeline at FDA




**If meeting is scheduled before day 75, written feedback must be given 5 days prior to meeting*

Traditional 510(k) Timeline at FDA




De Novo Timeline at FDA




**Note that based on FDA's internal timeline, the request for additional information will typically come by day 75*

IDE Timeline at FDA



Traditional PMA Timeline at FDA



**FDA review clock begins as of the date of receipt of the most recent submission or amendment that made the PMA complete and on which the FDA based its "Accepted" and "Filed" decisions.*

The background of the slide is a photograph of the United States Capitol building at night. The building is illuminated, with its iconic dome and classical architecture clearly visible. A large, white, thin-lined circle is superimposed over the center of the image, framing the text. The text is in a bold, white, sans-serif font.


Quality Management Considerations

Quality System Regulations


- All manufacturers of medical devices for distribution in the US have a Quality Management System, or QMS (often called a “Quality System” by FDA).
 - The QMS complies with Part 820 Quality System Regulation (QSR), which is similar to ISO 13485.
 - The QSR gives requirements for a range of quality activities – receiving goods, training personnel, handling complaints, etc.
 - Typically, each quality activity has a separate Standard Operating Procedure (SOP) in the manufacturer’s QMS.
- Besides complying with the QSR (Part 820), the QMS also complies with:
 - Part 7 Subpart C (Recalls)
 - Part 801 Labeling
 - Part 803 Medical Device Reporting
 - Part 806 Medical Devices, Reports of Corrections and Removals
 - Part 830 Unique Device Identification

Phased Implementation of QMS


Phase I: Initial Policies and Procedures



Phase II: Quality Manual and Intermediate Policies and Procedures



Phase III: Remaining Policies and Procedures for Commercial Distribution



The number of SOPs isn't dictated by FDA. It depends in part on the manufacturer's preferences, and on the type of device. For example, SaMDs require fewer SOPs than SiMDs.

A photograph of the U.S. Capitol building at night, illuminated by warm lights. A large white circle is superimposed over the center of the image, framing the text. The building's dome and classical architecture are visible against a dark sky.

Expedited Pathways

Breakthrough Device Designation

- FDA Goal: Encourage medical technology development and patient access under current pre-market review pathway
- Benefits:
 - Increased FDA interaction to facilitate development - “Sprint Discussion”
 - Expedited review of pre-market submission (510(k), De Novo, PMA)
 - Possible commercial/marketing benefit with FDA recognition of technology
- NOT:
 - Marketing authorization (still need to submit 510(k), De Novo, or PMA)
 - No finalized and implemented reimbursement benefit
- Requirements:
 - Criteria 1: provide for more effective treatment or diagnosis of life-threatening or irreversibly debilitating diseases or conditions
 - Demonstrate technical AND clinical success of device performance
 - Criteria 2: one or more of the following
 - A. represent breakthrough technologies
 - B. no approved or cleared alternatives exist
 - C. offer significant advantages over existing approved or cleared alternatives, including the potential, compared to existing approved alternatives, to reduce or eliminate the need for hospitalization, improve patient quality of life, facilitate patients’ ability to manage their own care (such as through self-directed personal assistance), or establish long-term clinical efficiencies; or
 - D. availability of which is in the best interest of patients

Safer Technologies Program (STeP)

- FDA Goal: Encourage medical technology development and patient access under current pre-market review pathway (like Breakthrough Device Program)
- Benefits:
 - Increased FDA interaction to facilitate development - “Sprint Discussion”
 - Expedited review of pre-market submission (510(k), De Novo, PMA)
- NOT:
 - Marketing authorization (still need to submit 510(k), De Novo, or PMA)
- Requirements:
 - Criteria 1: should not be eligible for the Breakthrough Devices Program due to the less serious nature of the disease or condition treated, diagnosed, or prevented by the device
 - Demonstrate technical AND clinical success of device performance
 - Criteria 2: should be reasonably expected to significantly improve the benefit-risk profile of a treatment or diagnostic through substantial safety innovations that provide for one or more of the following:
 - A. a reduction in the occurrence of a known serious adverse event,
 - B. a reduction in the occurrence of a known device failure mode,
 - C. a reduction in the occurrence of a known use-related hazard or use error, or
 - D. an improvement in the safety of another device or intervention.

Let's Pause to Ask Ourselves...

What pathway(s) should I be considering?

A photograph of the U.S. Capitol building at night, illuminated by warm lights. The building's iconic dome and neoclassical architecture are visible. In the foreground, a statue on a pedestal stands in front of the main entrance. The entire image is overlaid with a large, thin white circle. The text "Filing Logistics" is centered within this circle in a bold, white, sans-serif font.


Filing Logistics

MDUFA IV User Fees as of Oct 1, 2022

Application Type	Standard Fee	Small Business Fee†
510(k)	\$19,870	\$4,967
513(g)	\$5,961	\$2,980
PMA, PDP, PMR, BLA	\$441,547	\$110,387
De Novo Classification Request	\$132,464	\$33,116
Panel-track Supplement	\$353,238	\$88,309
180-Day Supplement	\$66,232	\$16,558
Real-Time Supplement	\$30,908	\$7,727
BLA Efficacy Supplement	\$441,547	\$110,387
30-Day Notice	\$7,065	\$3,532
Annual Fee for Periodic Reporting on a Class III device (PMAs,PDPs)	\$15,454	\$3,864

Small Business Application

- A small business is defined as a business, including its affiliates, whose gross receipts and sales are less than \$100 million for the most recent tax year.
- To apply, you must gather your tax documentation, obtain an organization ID, submit the application to FDA (Form 3602).
- This process could take several months, therefore, we recommend starting this now if you are interested in pursuing.
- [Reduced Medical Device User Fees: Small Business Determination \(SBD\) Program | FDA](#)



Thank you for your time and attention!


acadotte@mcra.com

Northwestern Medicine Healthcare AI Forum

MCRA AI & Imaging Center
Nima Akhlaghi, PhD.


February 9, 2024

Human Conditions




>10,000 Conditions Needing AI Software

FDA Software Approvals Since 1995




>600 over 25 Years
2022/2023 over >160 Approvals




Assessment by American College of Radiology:



Majority of FDA approvals are in therapeutic areas such Brain, Breast, Lung and Chest.




Retrospective Studies
Read by experts without impact on patient care




Ground Truthing

- Establish reference standard based on Standard of Care (SOC) to compare to device output.




Stand Alone Testing

- Comparing device output to reference standard



Multi-Reader, Multi-Case

- Comparing Aided vs Un-Aided performance against established reference standard



Thank you